



	Penetration panel	AQL Screened Room 6 was commissioned in 2008 and was designed with complex systems in mind. It is anechoically lined in accordance with the requirements of RTCA DO-160 and MIL STD 461D, allowing both very sensitive emissions testing and high power susceptibility testing up to 40GHz. The chamber is fully equipped for both single and three phase 50Hz, 60Hz, 400Hz and 800Hz AC supplies and high current DC supplies and features the common penetration panel arrangements used throughout the facility. As with SR 8, this room is independent and is ideal for secure/sensitive equipment that may require locking away from the main test lab.
	Room access	
	Ground plane	
	Anechoic lining	
	Supplies	
	Test / Support area	